

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

RMS
24-9767

Date: 03/05/2024

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:41
INTERNAL STANDARD	VERIFIED	12:41
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:41
BLANK TEST	0.00	12:42
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:42
BLANK TEST	0.00	12:43
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:44
BLANK TEST	0.00	12:44

