

SOLUTION CHANGE

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 01/02/2024

RMS
24-192

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:13
INTERNAL STANDARD	VERIFIED	11:13
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	11:13
BLANK TEST	0.00	11:14
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	11:14
BLANK TEST	0.00	11:15
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	11:15
BLANK TEST	0.00	11:16

