

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 02/21/2024

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:29
INTERNAL STANDARD	VERIFIED	07:29
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	07:29
BLANK TEST	0.00	07:30
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	07:30
BLANK TEST	0.00	07:31
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	07:31
BLANK TEST	0.00	07:32

LM
21-7700

