

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

RMS
24-9767

Date: 03/05/2024

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36882
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:51
INTERNAL STANDARD	VERIFIED	11:51
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	11:51
BLANK TEST	0.00	11:52
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	11:52
BLANK TEST	0.00	11:53
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	11:53
BLANK TEST	0.00	11:54

