

SOLUTION CHANGE

UTICA CITY  
 POLICE DEPT  
 DATAMASTER dmt: 100282

Date: 09/01/2021

OPERATOR NAME: *RMS*  
 MICHAEL J CLIVE  
 Permit #: 36682  
 Operator Agency/Dept: UTICA CITY PD  
 Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:09
INTERNAL STANDARD	VERIFIED	14:10
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	14:10
BLANK TEST	0.00	14:11
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	14:11
BLANK TEST	0.00	14:12
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	14:12
BLANK TEST	0.00	14:13

