

SOLUTION CHANGE

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

RMS
22-559

Date: 01/05/2022

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 366682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:21
INTERNAL STANDARD	VERIFIED	14:21
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:21
BLANK TEST	0.00	14:22
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:23
BLANK TEST	0.00	14:23
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:24
BLANK TEST	0.00	14:24

