

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 02/02/2021

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 20500

RMS
21-4012

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:03
INTERNAL STANDARD	VERIFIED	15:03
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	15:04
BLANK TEST	0.00	15:05
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	15:05
BLANK TEST	0.00	15:06
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	15:06
BLANK TEST	0.00	15:07

