

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

RMS
21-5752

Date: 02/16/2021

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:49
INTERNAL STANDARD	VERIFIED	12:49
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:50
BLANK TEST	0.00	12:50
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:51
BLANK TEST	0.00	12:51
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:52
BLANK TEST	0.00	12:53

