

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 04/20/2021

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 21020

RMS
21-14330

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:48
INTERNAL STANDARD	VERIFIED	13:48
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:48
BLANK TEST	0.00	13:49
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:49
BLANK TEST	0.00	13:50
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:50
BLANK TEST	0.00	13:51

