

SUPERVISOR TEST

UTICA CITY  
 POLICE DEPT  
 DATAMASTER dmt: 100282

*RMS*  
*21-15374*

Date: 04/27/2021

OPERATOR NAME:  
 MICHAEL J CLIVE  
 Permit #: 36682  
 Operator Agency/Dept: UTICA CITY PD  
 Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:10
INTERNAL STANDARD	VERIFIED	11:10
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	11:10
BLANK TEST	0.00	11:11
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	11:11
BLANK TEST	0.00	11:12
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	11:12
BLANK TEST	0.00	11:13

