

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 100282

ZMS
21-17440

Date: 05/11/2021

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:22
INTERNAL STANDARD	VERIFIED	15:22
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	15:23
BLANK TEST	0.00	15:23
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	15:24
BLANK TEST	0.00	15:25
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	15:25
BLANK TEST	0.00	15:26

