

SUPERVISOR TEST

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 100282

Date: 05/25/2021

*RMS  
21-19871*

OPERATOR NAME:  
MICHAEL J CLIVE  
Permit #: 36682  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:37
INTERNAL STANDARD	VERIFIED	12:37
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:38
BLANK TEST	0.00	12:38
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:39
BLANK TEST	0.00	12:40
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:40
BLANK TEST	0.00	12:41

