

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 09/10/2021

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:58
INTERNAL STANDARD	VERIFIED	12:58
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:58
BLANK TEST	0.00	12:59
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:59
BLANK TEST	0.00	13:00
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:00
BLANK TEST	0.00	13:01

2ms
21-31391

