

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 100282

RMS
21-30321

Date: 09/14/2021

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:55
INTERNAL STANDARD	VERIFIED	12:55
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:56
BLANK TEST	0.00	12:56
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:57
BLANK TEST	0.00	12:58
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:58
BLANK TEST	0.00	12:59

