

SUPERVISOR TEST

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 107308

Date: 04/27/2021

OPERATOR NAME:  
MICHAEL J CLIVE  
Permit #: 36682  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 21020

*RMS  
21-15374*

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:16
INTERNAL STANDARD	VERIFIED	11:16
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	11:16
BLANK TEST	0.00	11:17
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	11:18
BLANK TEST	0.00	11:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	11:19
BLANK TEST	0.00	11:19

