

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 107308

RM
 21-29379

Date: 07/27/2021

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:33
INTERNAL STANDARD	VERIFIED	11:33
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	11:34
BLANK TEST	0.00	11:34
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	11:35
BLANK TEST	0.00	11:35
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	11:36
BLANK TEST	0.00	11:37

