

SUPERVISOR TEST

UTICA CITY  
 POLICE DEPT  
 DATAMASTER dmt: 107308

*TCMS*  
*21-10/18/21*

Date: 10/12/2021

OPERATOR NAME:  
 MICHAEL J CLIVE  
 Permit #: 36682  
 Operator Agency/Dept: UTICA CITY PD  
 Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:00
INTERNAL STANDARD	VERIFIED	12:00
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:01
BLANK TEST	0.00	12:02
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:02
BLANK TEST	0.00	12:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:03
BLANK TEST	0.00	12:04

