

SOLUTION CHANGE

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 107308

Date: 11/01/2021

*EMS  
21-43292*

OPERATOR NAME:  
MICHAEL J CLIVE  
Permit #: 36682  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:58
INTERNAL STANDARD	VERIFIED	12:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:58
BLANK TEST	0.00	12:59
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:59
BLANK TEST	0.00	13:00
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:00
BLANK TEST	0.00	13:01

