

SOLUTION CHANGE

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 100282

RMS
ZZ-16003

Date: 05/02/2022

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:18
INTERNAL STANDARD	VERIFIED	14:18
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:18
BLANK TEST	0.00	14:19
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:19
BLANK TEST	0.00	14:20
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:21
BLANK TEST	0.00	14:21

