

SOLUTION CHANGE

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 100282

Date: 11/01/2022

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 22260

RMJ
22-44/60

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:35
INTERNAL STANDARD	VERIFIED	12:35
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:36
BLANK TEST	0.00	12:36
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:37
BLANK TEST	0.00	12:38
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:38
BLANK TEST	0.00	12:39

