

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 100282

RM
22-15118

Date: 04/26/2022

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:03
INTERNAL STANDARD	VERIFIED	12:04
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:04
BLANK TEST	0.00	12:05
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:05
BLANK TEST	0.00	12:06
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:06
BLANK TEST	0.00	12:07

