

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 06/21/2022

RM
22-21957

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:02
INTERNAL STANDARD	VERIFIED	13:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:02
BLANK TEST	0.00	13:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	13:03
BLANK TEST	0.00	13:04
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:04
BLANK TEST	0.00	13:05

