

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 07/05/2022

RMS
22-26077

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:09
INTERNAL STANDARD	VERIFIED	13:09
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:09
BLANK TEST	0.00	13:10
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:11
BLANK TEST	0.00	13:11
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:12
BLANK TEST	0.00	13:13

