

SOLUTION CHANGE

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 107308

Date: 09/30/2022

*RMS*  
*22-39426*

OPERATOR NAME:  
MICHAEL J CLIVE  
Permit #: 36682  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:11
INTERNAL STANDARD	VERIFIED	16:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	16:12
BLANK TEST	0.00	16:13
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	16:13
BLANK TEST	0.00	16:14
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	16:14
BLANK TEST	0.00	16:15

