

SUPERVISOR TEST

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 107308

Date: 05/24/2022

*RMS*  
*22-19468*

OPERATOR NAME:  
MICHAEL J CLIVE  
Permit #: 36682  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:19
INTERNAL STANDARD	VERIFIED	12:19
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:19
BLANK TEST	0.00	12:20
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:20
BLANK TEST	0.00	12:21
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:21
BLANK TEST	0.00	12:22

