

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 08/16/2022

RMS
22-32700

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:36
INTERNAL STANDARD	VERIFIED	13:36
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:36
BLANK TEST	0.00	13:37
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:37
BLANK TEST	0.00	13:38
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:38
BLANK TEST	0.00	13:39

