

SUPERVISOR TEST

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 107308

Date: 10/11/2022

*RMS*  
*22-4/1021*

OPERATOR NAME:  
MICHAEL J CLIVE  
Permit #: 36682  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:44
INTERNAL STANDARD	VERIFIED	14:44
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	14:45
BLANK TEST	0.00	14:46
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	14:46
BLANK TEST	0.00	14:47
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	14:47
BLANK TEST	0.00	14:48

