

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 12/16/2021

ZMS
21-49741

OPERATOR NAME:
CHRISTOPHER S JENNINGS
Permit #: 32846
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:03
INTERNAL STANDARD	VERIFIED	09:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	09:04
BLANK TEST	0.00	09:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	09:05
BLANK TEST	0.00	09:06
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	09:06
BLANK TEST	0.00	09:07

