

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 100282

Date: 05/02/2023

RMS
23-19475

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:59
INTERNAL STANDARD	VERIFIED	11:59
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	11:59
BLANK TEST	0.00	12:00
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:00
BLANK TEST	0.00	12:01
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:01
BLANK TEST	0.00	12:02

