

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 100282

Date: 06/27/2023

RMS
23-28922

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:44
INTERNAL STANDARD	VERIFIED	11:44
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	11:45
BLANK TEST	0.00	11:45
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	11:46
BLANK TEST	0.00	11:47
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	11:47
BLANK TEST	0.00	11:48

