

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 100282

RMS
23-324/15

Date: 07/18/2023

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:53
INTERNAL STANDARD	VERIFIED	10:53
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	10:54
BLANK TEST	0.00	10:54
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	10:55
BLANK TEST	0.00	10:56
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	10:56
BLANK TEST	0.00	10:57

