

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 08/29/2023

RW
25-394/91

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:33
INTERNAL STANDARD	VERIFIED	12:33
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:33
BLANK TEST	0.00	12:34
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:34
BLANK TEST	0.00	12:35
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:35
BLANK TEST	0.00	12:36

