

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

LMJ
23-47113

Date: 10/17/2023

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:27
INTERNAL STANDARD	VERIFIED	09:27
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	09:28
BLANK TEST	0.00	09:29
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	09:29
BLANK TEST	0.00	09:30
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	09:30
BLANK TEST	0.00	09:31

