

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 12/05/2023

RM
23-54567

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:05
INTERNAL STANDARD	VERIFIED	12:05
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:05
BLANK TEST	0.00	12:06
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:06
BLANK TEST	0.00	12:07
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:07
BLANK TEST	0.00	12:08

