

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 01/24/2023

RMS
23-3668

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:33
INTERNAL STANDARD	VERIFIED	13:33
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:33
BLANK TEST	0.00	13:34
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:34
BLANK TEST	0.00	13:35
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	13:36
BLANK TEST	0.00	13:36

