

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 107308

Date: 03/28/2023

RMS
23-13458

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 22420

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:42
INTERNAL STANDARD	VERIFIED	13:42
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:42
BLANK TEST	0.00	13:43
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:43
BLANK TEST	0.00	13:44
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:45
BLANK TEST	0.00	13:45

