

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

RMS
23-19475

Date: 05/02/2023

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:55
INTERNAL STANDARD	VERIFIED	11:55
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	11:56
BLANK TEST	0.00	11:56
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	11:57
BLANK TEST	0.00	11:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	11:58
BLANK TEST	0.00	11:59

