

SUPERVISOR TEST

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 107308

Date: 05/16/2023

*RMS*  
*23-22080*

OPERATOR NAME:  
MICHAEL J CLIVE  
Permit #: 36682  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:33
INTERNAL STANDARD	VERIFIED	10:33
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	10:33
BLANK TEST	0.00	10:34
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	10:34
BLANK TEST	0.00	10:35
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	10:35
BLANK TEST	0.00	10:36

