

SOLUTION CHANGE

UTICA CITY  
 POLICE DEPT  
 DATAMASTER dmt: 107308

*RMS*  
*23-29869*

Date: 07/03/2023

OPERATOR NAME:  
 MICHAEL J CLIVE  
 Permit #: 36682  
 Operator Agency/Dept: UTICA CITY PD  
 Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:19
INTERNAL STANDARD	VERIFIED	12:19
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:19
BLANK TEST	0.00	12:20
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:20
BLANK TEST	0.00	12:21
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:21
BLANK TEST	0.00	12:22

