

SUPERVISOR TEST

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 107308

Date: 07/05/2023

*RMS*  
*23-30229*

OPERATOR NAME:  
MICHAEL J CLIVE  
Permit #: 36682  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:13
INTERNAL STANDARD	VERIFIED	14:13
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	14:14
BLANK TEST	0.00	14:15
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	14:15
BLANK TEST	0.00	14:16
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	14:16
BLANK TEST	0.00	14:17

