

SUPERVISOR TEST

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 107308

Date: 09/19/2023

*RMS*  
*23-4/2835*

OPERATOR NAME:  
MICHAEL J CLIVE  
Permit #: 36682  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:42
INTERNAL STANDARD	VERIFIED	10:42
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	10:42
BLANK TEST	0.00	10:43
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	10:43
BLANK TEST	0.00	10:44
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	10:45
BLANK TEST	0.00	10:45

