

SUPERVISOR TEST

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 107308

Date: 10/17/2023

*KMS*  
*23-47113*

OPERATOR NAME:  
MICHAEL J CLIVE  
Permit #: 36682  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:36
INTERNAL STANDARD	VERIFIED	09:36
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	09:37
BLANK TEST	0.00	09:37
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	09:38
BLANK TEST	0.00	09:38
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	09:39
BLANK TEST	0.00	09:40

