

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 100282

Date: 11/05/2024

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:46
INTERNAL STANDARD	VERIFIED	10:46
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	10:46
BLANK TEST	0.00	10:47
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	10:47
BLANK TEST	0.00	10:48
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	10:48
BLANK TEST	0.00	10:49

