

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

*12MS
25-211*

Date: 12/31/2024

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:29
INTERNAL STANDARD	VERIFIED	11:29
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	11:29
BLANK TEST	0.00	11:30
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	11:31
BLANK TEST	0.00	11:31
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	11:32
BLANK TEST	0.00	11:32

