

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 01/09/2024

LMS
24-1289

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:48
INTERNAL STANDARD	VERIFIED	09:48
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	09:49
BLANK TEST	0.00	09:50
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	09:50
BLANK TEST	0.00	09:51
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	09:51
BLANK TEST	0.00	09:52

