

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 05/14/2024

AMS
24-21111

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:33
INTERNAL STANDARD	VERIFIED	07:34
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	07:34
BLANK TEST	0.00	07:35
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	07:35
BLANK TEST	0.00	07:36
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	07:36
BLANK TEST	0.00	07:37

