

SOLUTION CHANGE

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 100282

RMS
24-34205

Date: 08/02/2024

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:12
INTERNAL STANDARD	VERIFIED	10:12
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	10:12
BLANK TEST	0.00	10:13
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	10:13
BLANK TEST	0.00	10:14
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	10:14
BLANK TEST	0.00	10:15

