

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 08/20/2024

RMS
24-3763

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:53
INTERNAL STANDARD	VERIFIED	11:53
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	11:53
BLANK TEST	0.00	11:54
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	11:54
BLANK TEST	0.00	11:55
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	11:55
BLANK TEST	0.00	11:56

