

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 08/27/2024

12 Ms
261-38186

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:43
INTERNAL STANDARD	VERIFIED	07:43
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	07:43
BLANK TEST	0.00	07:44
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	07:44
BLANK TEST	0.00	07:45
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	07:45
BLANK TEST	0.00	07:46

