

SUPERVISOR TEST

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 100282

Date: 09/10/2024

*RMS*  
*24-404/73*

OPERATOR NAME:  
MICHAEL J CLIVE  
Permit #: 36682  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:35
INTERNAL STANDARD	VERIFIED	12:35
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:36
BLANK TEST	0.00	12:36
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:37
BLANK TEST	0.00	12:38
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:38
BLANK TEST	0.00	12:39

